



info@scigenus.com or www.scigenus.com

CERTIFICATE OF ANALYSIS

Product Description: ZINC BUFFERED 10% FORMALIN

Scigen Product #: 1030Z, 1060Z, 1120Z, 1180Z, 1480Z, 2005Z, 6064Z, 5041

Test Code	Test Element	Result	Specification
3300	Appearance	Pass	Clear Liquid
1100	pH	Pass	5.1-6.0
900	Assay, Formalin	10.7%	10.0 - 11.0%
	Lot No.	4635	
	Exp. Date	Jan-27	
	Meets Specification:	Yes	
	Approval Date:	1/25/2023	

CERTIFIED WASTE DISPOSAL:



IT IS RECOMMENDED TO TREAT FORMALIN WASTE WITH **NEUTRALEX®**, A GOVERNMENT CERTIFIED SOLUTION RESULTING IN DRAIN SAFE WASTE. **NEUTRALEX®** PROVIDES COMPLETE NEUTRALIZATION IN 15 MINUTES WITH NO pH ADJUSTEMENTS NEEDED.

AFTER TREATMENT, **NEUTRALEX®** IS CERTIFIED TO PRODUCE A NON-TOXIC SOLUTION THAT CAN BE DISPOSED INTO THE SEWER SYSTEM WITHOUT FEAR OF ENVIRONMENTAL DAMAGE

ACCIDENTAL SPILLS AND RELEASE MEASURES:

- LARGE SPILLS (OVER 40 mL): TREAT IMMEDIATELY WITH NEUTRA-FORM®
- SMALL SPILLS (LESS THAN 40mL): ABSORB AND NEUTRALIZE WITH NEUTRA-WIPES®.

REQUEST INFORMATION: EMAIL INFO@SCIGENUS.COM OR CALL (310) 324-6576

- 4046 | NEUTRALEX® STARTER KIT
- 4047 | NEUTRALEX® CASE OF 16 PREMEASURED PACKETS
- 4304 | NEUTRA-FORM® BULK SPILL NEUTRALIZING AGENT
- 4278 | NEUTRA-WIPES® NEUTRALIZE SMALL FORMALIN SPILLS
- 4295 | NEUTRA-PADS® NEUTRALIZING GROSSING PADS
- 4050 | NEUTRALEX® ALDEHYDE TEST KIT